



Analytical[®]
Technologies Limited



WATER BATH

SPECIFICATION

- Temperature Sensitivity: \pm °C 0.5 or better
- Spatial Deviation In Temperature: \pm °C 0.5 or better
- Readability °C : 0.1
- Temperature range °C: 5°C above ambient to 90°C
- Temperature Controller: Solid State digital Controller/PID
- Display : LED/LCD



Analytical[®]
Technologies Limited

Analytical Group Companies
Analytical Technologies Limited
Analytical Bio-Technologies
Analytical Bio- Med
HPLC Solutions

Branch Office / Distributor

Corporate & Regd. Office :

Analytical House, # E67 & E68. Ravi Park, Opp., Sainath Park, Vasna Road, Baroda, Gujarat - 390 015. INDIA
Tele : 0265-2253620 / 2252839 Fax : 0265-2254395 E-mail : info@hplctechnologies.com / info@ais-india.com
Web : www.ais-india.com / www.analyticalgroup.net